



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/924,021	08/07/2001	Tomoaki Ito	08228/017001	3025

22511 7590 06/29/2004

OSHA & MAY L.L.P.  
1221 MCKINNEY STREET  
HOUSTON, TX 77010

EXAMINER
----------

SIEFKE, SAMUEL P

ART UNIT	PAPER NUMBER
----------	--------------

1743

DATE MAILED: 06/29/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

09/924,021

Applicant(s)

ITO ET AL.

Examiner

Samuel P Siefke

Art Unit

1743

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 02 April 2004.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-6, 14-20 and 23 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-6, 14-20, 23 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)   | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## DETAILED ACTION

### *Response to Arguments*

Applicant's arguments filed 4/2/4 have been fully considered but they are not persuasive. Applicant argues, "The pH difference is infallible indication of the exhaustion of the anion exchanger to be monitored. In this respect it does not matter whether weak or strong anions break through first, for example, traces of carbonic acid or silicic acid which breaks through are easily detected, "(col. 1, l. 66-col. 8, l. 4). The above clearly indicates that Ammer simply measures the ion concentration irrespective of the type of acid ion." This rejection has been withdrawn but new rejection has been made.

### *Claim Rejections - 35 USC § 112*

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims **1-6, 14-20** and **23** rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. The amendment on 4/2/04 has new matter. There is no basis in the specification for claiming "**selectively** measuring an inorganic

Art Unit: 1743

carbonic acid concentration in an outlet water of an ion exchange resin vessel filled with at least an anion exchange resin, **thereby avoiding influences from other acid ions**"

The specification states on page 23 and 24, "the gas permeation membrane used here allows the carbon dioxide and other gases permeate though, but does not allow the ionic constituents and organic materials." Then goes on to say, "**When there is no other ionic constituent present**, the carbonic acid concentration (carbon dioxide concentration) and the electric conductivity of the water dissociation a function represented by equilibrium constant which is a function of the temperature...It is clear from the above that because, in the conductivity sensor, the influence of any ionic constituents other than carbon dioxide is averted by the gas permeation membrane, the inorganic carbonic acid concentration can be measured accurately even for sample water having other ion constituents such as chloride and sulfuric acid ions." To claim "selectively" measuring an inorganic carbonic acid concentration in outlet water of an ion exchange resin where unknown sample are present, other gases can pass through the barrier for example. Since the specification only has basis for a gas permeation membrane and not a selective gas permeation membrane for carbon dioxide a new matter rejection is issued.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Samuel P Siefke whose telephone number is 571-272-1262. The examiner can normally be reached on M-F 7:00am-5:00pm.


Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on 571-272-1700. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Sam P. Siefke

A handwritten signature in black ink, consisting of a series of loops and curves, positioned below the name Sam P. Siefke.

June 28, 2004

A handwritten signature in black ink, appearing to read 'Jill Warden', positioned above the printed name and title.

Jill Warden  
Supervisory Patent Examiner  
Technology Center 1700